



See O-Rinse Machine Instruction Manual for meaning of symbols

O-Rinse Bottles and Assorted Spares. For unabridged instructions see the Instruction Manual supplied with your machine. A

- copy may be obtained from Altomed if necessary. These instructions are a summary. This product is intended for use with the MicroWash O-Rinse or AOI Quickrinse Machine.
- Clean before use and at times determined by your risk assessment.
- Wear the correct PPE (personal protective equipment) gloves, apron, sleeves and visor or protective eyewear) or as recommended in your Risk Assessment.
- Submerge the product into a suitable CE marked disinfectant such as Ruhof Virusolve for the time designated on the label. Use a syringe to draw disinfectant into any tubing.
- Rinse thoroughly in sterile or RO water, use a syringe to draw water through the tubes.
- This device does not contact the patient and as such can be decontaminated by disinfection. If required the bottle assembly can also withstand exposure to a standard autoclave cycle of 134°C to 137°C with a 3 to 3½ minute holding time. Ensure the bottle cap has all tubing connected and has been removed from the bottle.
- Do not put the bottle in a peel pouch or leave the cap on as this may cause it to collapse.
- Exposure to strong alkalines may cause discolouration, this will be cosmetic damage only.





5th June 2019

ALT 1081 Issue 02/0619

See Q-Rinse Machine Instruction Manual for meaning of symbols

Q-Rinse Bottles and Assorted Spares.

- For unabridged instructions see the Instruction Manual supplied with your machine. A copy may be obtained from Altomed if necessary. These instructions are a summary
- This product is intended for use with the MicroWash Q-Rinse or AOI Quickrinse Machine.
- Clean before use and at times determined by your risk assessment.

Altomed Ltd, 2 Witney Way, Tyne/Wear, NE35 9PE, UK.

- Wear the correct PPE (personal protective equipment) gloves, apron, sleeves and visor or protective eyewear) or as recommended in your Risk Assessment.
- Submerge the product into a suitable CE marked disinfectant such as Ruhof Virusolve for the time designated on the label. Use a syringe to draw disinfectant into any tubing.
- Rinse thoroughly in sterile or RO water, use a syringe to draw water through the tubes.
- This device does not contact the patient and as such can be decontaminated by disinfection. If required the bottle assembly can also withstand exposure to a standard autoclave cycle of 134°C to 137°C with a 3 to 3½ minute holding time. Ensure the bottle cap has all tubing connected and has been removed from the bottle.
- Do not put the bottle in a peel pouch or leave the cap on as this may cause it to collapse. Exposure to strong alkalines may cause discolouration, this will be cosmetic damage only.



ALT I081 Issue 02/0619 5th June 2019





See O-Rinse Machine Instruction Manual for meaning of symbols

Q-Rinse Bottles and Assorted Spares.

- For unabridged instructions see the Instruction Manual supplied with your machine. A copy may be obtained from Altomed if necessary. These instructions are a summary.
- This product is intended for use with the MicroWash Q-Rinse or AOI Quickrinse Machine.
- Clean before use and at times determined by your risk assessment.

Altomed Ltd, 2 Witney Way, Tyne/Wear, NE35 9PE, UK.

- Wear the correct PPE (personal protective equipment) gloves, aprop. sleeves and visor or protective evewear) or as recommended in your Risk Assessment.
- Submerge the product into a suitable CE marked disinfectant such as Ruhof Virusolve for the time designated on the label. Use a syringe to draw disinfectant into any tubing.
- Rinse thoroughly in sterile or RO water, use a syringe to draw water through the tubes
- This device does not contact the patient and as such can be decontaminated by disinfection. If required the bottle assembly can also withstand exposure to a standard autoclave cycle of 134°C to 137°C with a 3 to 3½ minute holding time. Ensure the bottle cap has all tubing connected and has been removed from the bottle.
- Do not put the bottle in a peel pouch or leave the cap on as this may cause it to collapse. Exposure to strong alkalines may cause discolouration, this will be cosmetic damage only.
- ALT I081 Issue 02/0619 Altomed Ltd, 2 Witney Way, Tyne/Wear, NE35 9PE, UK.






ALT I081 Issue 02/0619

5th June 2019

See Q-Rinse Machine Instruction Manual for meaning of symbols

Q-Rinse Bottles and Assorted Spares.

- For unabridged instructions see the Instruction Manual supplied with your machine. A copy may be obtained from Altomed if necessary. These instructions are a summary.
- This product is intended for use with the MicroWash Q-Rinse or AOI Quickrinse Machine.
- Clean before use and at times determined by your risk assessment.
- Wear the correct PPE (personal protective equipment) gloves, apron, sleeves and visor or protective eyewear) or as recommended in your Risk Assessment.
- Submerge the product into a suitable CE marked disinfectant such as Ruhof Virusolve for the time designated on the label. Use a syringe to draw disinfectant into any tubing.
- Rinse thoroughly in sterile or RO water, use a syringe to draw water through the tubes.
- This device does not contact the patient and as such can be decontaminated by disinfection. If required the bottle assembly can also withstand exposure to a standard autoclave cycle of 134°C to 137°C with a 3 to 3½ minute holding time. Ensure the bottle cap has all tubing connected and has been removed from the bottle.
- Do not put the bottle in a peel pouch or leave the cap on as this may cause it to collapse.
- Exposure to strong alkalines may cause discolouration, this will be cosmetic damage only.



altomed



See O-Rinse Machine Instruction Manual for meaning of symbols

Q-Rinse Bottles and Assorted Spares.

- For unabridged instructions see the Instruction Manual supplied with your machine. A copy may be obtained from Altomed if necessary. These instructions are a summary.
- 2. This product is intended for use with the MicroWash O-Rinse or AOI Quickrinse Machine.
- Clean before use and at times determined by your risk assessment. 3
- Wear the correct PPE (personal protective equipment) gloves, aprop. sleeves and visor or Δ protective evewear) or as recommended in your Risk Assessment.
- 5 Submerge the product into a suitable CE marked disinfectant such as Ruhof Virusolve for the time designated on the label. Use a syringe to draw disinfectant into any tubing. 6
- Rinse thoroughly in sterile or RO water, use a syringe to draw water through the tubes.
- This device does not contact the patient and as such can be decontaminated by 7. disinfection. If required the bottle assembly can also withstand exposure to a standard autoclave cycle of 134°C to 137°C with a 3 to 3½ minute holding time. Ensure the bottle cap has all tubing connected and has been removed from the bottle.
- Do not put the bottle in a peel pouch or leave the cap on as this may cause it to collapse.
- Exposure to strong alkalines may cause discolouration, this will be cosmetic damage only.



ALT 1081 Issue 02/0619 5th June 2019

li MD

See Q-Rinse Machine Instruction Manual for meaning of symbols

Q-Rinse Bottles and Assorted Spares.

- For unabridged instructions see the Instruction Manual supplied with your machine. A copy may be obtained from Altomed if necessary. These instructions are a summary.
- 2 This product is intended for use with the MicroWash Q-Rinse or AOI Quickrinse Machine.
- 3. Clean before use and at times determined by your risk assessment.

Altomed Ltd, 2 Witney Way, Tyne/Wear, NE35 9PE, UK.

- 4. Wear the correct PPE (personal protective equipment) gloves, apron, sleeves and visor or protective eyewear) or as recommended in your Risk Assessment.
- Submerge the product into a suitable CE marked disinfectant such as Ruhof Virusolve for the time designated on the label. Use a syringe to draw disinfectant into any tubing.
- 6. Rinse thoroughly in sterile or RO water, use a syringe to draw water through the tubes.
- This device does not contact the patient and as such can be decontaminated by disinfection. If required the bottle assembly can also withstand exposure to a standard autoclave cycle of 134°C to 137°C with a 3 to 3½ minute holding time. Ensure the bottle cap has all tubing connected and has been removed from the bottle.
- Do not put the bottle in a peel pouch or leave the cap on as this may cause it to collapse. 8.
- Exposure to strong alkalines may cause discolouration, this will be cosmetic damage only.





ALT 1081 Issue 02/0619

5th June 2019

See O-Rinse Machine Instruction Manual for meaning of symbols Q-Rinse Bottles and Assorted Spares.

- For unabridged instructions see the Instruction Manual supplied with your machine. A copy may be obtained from Altomed if necessary. These instructions are a summary.
- 2. This product is intended for use with the MicroWash Q-Rinse or AOI Quickrinse Machine.
- Clean before use and at times determined by your risk assessment. 3.
- Wear the correct PPE (personal protective equipment) gloves, apron. sleeves and visor or 4. protective evewear) or as recommended in your Risk Assessment.
- 5. Submerge the product into a suitable CE marked disinfectant such as Ruhof Virusolve for the time designated on the label. Use a syringe to draw disinfectant into any tubing.
- 6 Rinse thoroughly in sterile or RO water, use a syringe to draw water through the tubes.
- This device does not contact the patient and as such can be decontaminated by disinfection. If required the bottle assembly can also withstand exposure to a standard autoclave cycle of 134°C to 137°C with a 3 to 3½ minute holding time. Ensure the bottle cap has all tubing connected and has been removed from the bottle.
- Do not put the bottle in a peel pouch or leave the cap on as this may cause it to collapse.
- Exposure to strong alkalines may cause discolouration, this will be cosmetic damage only.



Altomed Ltd, 2 Witney Way, Tyne/Wear, NE35 9PE, UK. Tel: 0191 519 0111 - E. Mail: admin@altomed.com

ALT 1081 Issue 02/0619 5th June 2019





for meaning of symbols

Q-Rinse Bottles and Assorted Spares.

- For unabridged instructions see the Instruction Manual supplied with your machine. A copy may be obtained from Altomed if necessary. These instructions are a summary.
- 2. This product is intended for use with the MicroWash Q-Rinse or AOI Quickrinse Machine
- 3. Clean before use and at times determined by your risk assessment.
- 4. Wear the correct PPE (personal protective equipment) gloves, apron, sleeves and visor or protective eyewear) or as recommended in your Risk Assessment.
- 5. Submerge the product into a suitable CE marked disinfectant such as Ruhof Virusolve for the time designated on the label. Use a syringe to draw disinfectant into any tubing.
- Rinse thoroughly in sterile or RO water, use a syringe to draw water through the tubes.
- This device does not contact the patient and as such can be decontaminated by disinfection. If required the bottle assembly can also withstand exposure to a standard autoclave cycle of 134°C to 137°C with a 3 to 3½ minute holding time. Ensure the bottle cap has all tubing connected and has been removed from the bottle.
- Do not put the bottle in a peel pouch or leave the cap on as this may cause it to collapse.
- Exposure to strong alkalines may cause discolouration, this will be cosmetic damage only.



Altomed Ltd, 2 Witney Way, Tyne/Wear, NE35 9PE, UK. Tel: 0191 519 0111 - E. Mail: admin@altomed.com